

PROCEDURE QP-02, "CONTROL OF RECORDS", PARAGRAPH 8.0.

SUPPLIER SHALL USE CALIBRATED INSPECTION AND TEST EQUIPMENT TO VERIFY

PRODUCT AS REQUIRED, AND MAINTAINED IN ACCORDANCE WITH ANSI/NCSL

Z540.1 AND ISO 10012. [REV. 04/02/08]